

# FLOTEC INC

Web Site: <http://www.floteco2.com>

7625 West New York Street 317-273-6960  
Indianapolis, Indiana 46214-4911

Fax 317-273-6979

Order desk: 800-401-1723

e-mail: [flotec@floteco2.com](mailto:flotec@floteco2.com)

Order desk Fax 800-515-9254

The FLOTEC Australian Standard SIS oxygen Flowmeter is designed to provide a constant flow of oxygen at different inlet pressure. In addition, the FLOTEC Australian Standard SIS Flowmeter can simultaneously deliver a measured flow through the twelve-position selector. FLOTEC Australian Standard SIS oxygen Flowmeters have a metal inlet filter and rotor filter body and are available in a complete range of Australian SIS Nut and Nipple or 1/8" Female pipe inlet connections and Hose Barb or DISS fitting valve outlet connections. FLOTEC products contain no latex or latex byproducts.

SPECIFICATIONS:

Operating Range: 17, 20, 24, 30, 50, 58, 70 PSI

Flow Capacity: Minimum: 30 MLM; Maximum: 60 LPM  
thru the 12 position selector

Temperature Range: -20 to 150F

Inlet Configuration: Australian Standard SIS Clear color  
Anodized Nut and Brass Bright DIP Nipple  
with G 1/4" BSPP/BSPF thread

Outlet Configuration: Hose Barb, clear color anodized

REPLACEMENT PARTS:

Inlet and outlet Connections

- 316L S.S. Sintered Filter disk
- 20 micron Sintered Stainless Steel Rotor filter body w/ lased shim-cup

Rotor Assembly

O-ring the 1 1/16-20 FM Body THD

O-ring for Australian STD SIS Nipple

INSTALLATION: Oxygen Flowmeter Sub- Assembly:

Carefully remove red cap from inlet and outlet port by twisting out. **Do not pull cap straight out.**

Check to insure filter disk is installed. Do not allow Teflon tape to cover first thread of filling. Use no more than two turns of tape.

INSTALLATION: FM body with FMW Cover: Be sure is properly positioned Flowmeter Sub-Assembly and Cover before torque.

INSTALLATION: CGA SIS Nut and Nippler: Mount the Flowmeter with Australian Std SIS Flowmeter Assy to a SIS STD cylinder valve. Be careful to avoid cross-threading the connection. To achieve a proper seal for threaded cylinder connections, which use O-Ring be sure the O-Ring is properly positioned before attaching the regulator to the valve.

INSTALLATION: Pipe Connector: Use only Teflon tape on the threads as a sealant.

WARNING: Some Medical gases can be dangerous. Large quantities of some medical gases can push the oxygen out of a room which can cause a person to faint and/or asphyxiate. Consult CGA standard # G-8.1 for other cautions.

PREVENTIVE MAINTENANCE: All Flowmeters should be tested and cleaned periodically to insure proper performance. The frequency of testing should be established according to usage, but it should be performed at least once per year to evaluate for damage, contamination, wear and performance.

LEAK TEST: Attach the Flowmeter to outlet medical gas system, plug all flowmeter outlets, turn the flow selector to "OFF" and slowly open outlet medical system valve. Apply a compatible leak test solution to all outlets, fittings, and Flowmeter body to check for bubbles. Tighten fittings as required to eliminate all external leaks. DO NOT over tighten threaded connections. Replace o-ring if required. Close outlet system valve.

WARNING: Disassembly, assembly, and testing of Flowmeters with Australian Standard SIS should be performed only by trained personnel. The work area must be free of hydrocarbon contaminants and residues because of the danger of spontaneous combustion when residues are exposed to gaseous oxygen.

## SAFETY WARNINGS:

1. The use of FLOTEC Australian Standard SIS oxygen flowmeters for gases other than compressed outlet oxygen of up to 70 psi is expressly prohibited by FLOTEC and the user must assume all liabilities.
2. Use no oil or grease.
3. Never administer oxidizing gases when smoking or when near an open flame.
4. Never use medical gases from a cylinder without reducing the pressure through a suitable regulator intended for that gas.
5. **Ensure that the threaded fitting on flowmeters and Nut are properly mated for the gas intended.** Never attempt to force an incompatible connection.
6. Never permit compressed oxygen to enter a flowmeter suddenly. Always open the valve slowly.
7. Fully open the valve when a flowmeter is attached and in use.
8. Never leave a medical gas system valve open with flowmeter attached when flowmeter is not in use.
9. Before a flowmeter is removed, fully close the medical gas system valve and release all residual gas pressure from within the flowmeter.
10. Never interchange flowmeters, hoses, or other equipment with similar equipment intended for use with other gases. Flowmeters and related fittings should never be handled with oily or greasy hands or gloves. Never hold hand over the outlet(s) to test for the presence of pressure.
11. Never use oxygen as a pressure medium to purge obstructed pipelines or equipment, to operate pneumatic tools, or to build up pressure in any tanks.
12. Do not stand in front of a flowmeter outlet when opening the medical gas system valve in case foreign particles are present which could cause a hazardous malfunction of the flowmeter.
13. Oxygen therapy may be critical treatment. The use of the flowmeter should be made in strict accordance with the prescription and instructions of a physician.
14. Secure cylinders to wall, stand, or cart in accordance with local fire codes.
15. Downstream equipment used in conjunction with the flowmeter should be equipped with suitable safety valves to prevent overpressurization and damage.
16. Oxygen must never be allowed to contact oil, grease or other petroleum-based substances. Therefore, do not use oil or grease on oxygen flowmeters, cylinders, valves or other related equipment. Do not use or store oxygen equipment near excessive heat (>125 F or 51.5 C) or open flame.
17. CAUTION: Do not use organic-based threaded sealants on any portion of the regulator. Use only Teflon threaded-sealing tape or oxygen service compound.
18. Never leave pressurized oxygen within a flowmeter. Always purge residual gases when not in use.
19. The Flowmeter knurled Australian STD SIS Nut must be tight to prevent O-ring extrusion.
20. The Flowmeter knurled Australian STD SIS Nut must NOT be twisted or turned when it is tight on the cylinder valve and/or under pressure to prevent O-ring abrasion.

Flotec Patent #'s: 4909476, 5785050, 6026854, 6082396, 6328280, 7089956, 4982760, 6148841, 7255127

# FLOTEC INC

7625 West New York Street  
Indianapolis, Indiana 46214-4911

317-273-6960  
Fax 317-273-6979

Web Site: <http://www.floteco2.com>  
E-mail: [flotec@floteco2.com](mailto:flotec@floteco2.com)

Order desk: 800-401-1723  
Order desk Fax 800-515-9254

---

## Declaration of Conformity

### Product identification

Product name : Regulators & Flow meters, Regulators with Flow meter.  
Brand : Flotec  
Cat. Number : RXXXX-XXXXX, FXXX-XXXXX, MXXXX-XXXXX, DXXXX-XXXXX  
Batch/Serial Nr. : Numbered Sequentially

### Manufacturer

Name : Flotec  
Address : 7625 West New York Street  
Indianapolis, IN 46214-4911  
Country : USA  
Representative : Brian Davidson  
Tel : +1 317-273-6960  
Fax : +1 317-273-6979

### Authorized Representative in Europe

Name : Emergo Europe  
Address : Molenstraat 15  
2513 BH  
The Hague  
Country : The Netherlands  
Tel : (31) (0) 70 345-8570  
Fax : (31) (0) 70 346-7299

### Notified Body

Name : UL International (UK) Ltd  
Address : Womersley House, The Guildway  
Old Portsmouth Road  
Country : UK-Guildford, Surrey, GU3 1LR  
I.D. Number : CE 0843

### Means of conformity

Flotec Inc. declares that the product listed has been classified as Class IIb - Annex IX, Rule 11 and is in conformity with the essential requirements and provisions of Council Directive: 93/42/EEC

### Signature

Place and date : Flotec Inc 9-22-03

Signature :



Name : Brian Davidson  
[Name and Signature of authorized person within organization]

# FLOTEC INC

7625 West New York Street  
Indianapolis, Indiana 46214-4911

317-273-6960  
Fax 317-273-6979

Web Site: <http://www.floteco2.com>  
E-mail: [flotec@floteco2.com](mailto:flotec@floteco2.com)

Order desk: 800-401-1723  
Order desk Fax 800-515-9254

---



## Authorized Representative

Name : Emergo Europe  
Address : Molenstraat 15  
2513 BH The Hague  
Country : The Netherlands  
Tel : (31) (0) 70 345-8570

---

## Symbols

### Symbols

### Definition

PSI  
KG/CM<sup>2</sup>  
BAR  
MPA  
HP  
LPM  
SN

POUNDS PER SQUARE INCH  
KILOGRAMS PER SQUARE CENTIMETER  
KILOGRAMS PER SQUARE CENTIMETER  
MEGA PASCAL  
HIGH PRESSURE PORT  
LITERS PER MINUTE  
SERIAL NUMBER



USE NO OIL



CONSULT INSTRUCTIONS FOR USE



AUTHORISED REPRESENTATIVE IN THE EUROPEAN COMMUNITY